



Blood Cell Storage, Inc.
4203 23rd Ave W,
Seattle WA, 98199 USA

Certificate of Compliance

Date of EtO Processing: 30 JUN 2021

Product Name: SAFE Sens[®] sv² sensor kit

Part Number: 400003

Lot Number: 2105031

Storage: Room temperature protected from light


EXPIRY: 29 FEB 2024

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- ✓ Mouse embryo toxicity Pass
- ✓ Biological Indicator Culturing Test Pass
- ✓ EtO Sterilized (SAL 10⁻⁶) Pass
- ✓ Precise pH interpretation, pHID A3929
(when used at 35°-38°C)





Steve Geelhood
Business Director

11 Feb 2022

Date

BCSI Internal Tracking: 2202023

400003.06

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSL
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN
PERMISSION FROM BCSL'S QUALITY ASSURANCE MANAGER.

