



Blood Cell Storage, Inc.
 123 NW 36th Street, Suite 220
 Seattle, WA 98107
 USA

Certificate of Compliance

Date of EtO Processing: 21 August 2020

Product Name: SAFE Sens[®] sv² sensor kit

Part Number: 400010

Lot Number: 1912015

Storage: Room temperature protected from light

EXPIRY: 30 April 2022

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- | | |
|--|--------------|
| ✓ Mouse embryo toxicity | Pass |
| ✓ Biological Indicator Culturing Test | Pass |
| ✓ EtO Sterilized (SAL 10 ⁻⁶) | Pass |
| ✓ Precise pH interpretation, encrypted pHID
(when used at 35°-38°C) | Y3GH7 |



Irena Dikovskiy
 Director of Quality

Date

BCSI Internal Tracking: 2101013

400010.04

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI.
 THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN
 PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

