



Certificate of Compliance

Date of EtO Processing: 21 August 2020

Product Name: sv² sensor kit

Part Number: 400019

Lot Number: 1912015

Storage: Room temperature protected from light

EXPIRY: 30 April 2022

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard and ISO 13485:2016 Medical Devices Quality Management System.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:
 - ✓ Mouse embryo toxicity Pass
 - ✓ Biological Indicator Culturing Test Pass
 - ✓ EtO Sterilized (SAL 10⁻⁶) Pass
 - ✓ Precise pH interpretation, encrypted pHID ^%@OM
(when used at 35°-38°C)



Irena Dikovskiy
Director of Quality

01 Dec 2020

Date

BCSI Reference: 2011022

Rev 400019.02

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI.
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN
PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

