SAFE Sens

Blood Cell Storage, Inc.
4203 23rd Ave W,
Seattle WA, 98199 USA

Certificate of Compliance

Date of EtO Processing: 30 June 2021

Product Name: SAFE Sens® sv^2 sensor kit

Part Number: 400003

Lot Number: 2104028

Storage: Room temperature protected from light

EXPIRY: 31 October 2023

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

✓ Mouse embryo toxicity	Pass
✓ Biological Indicator Culturing Test	Pass
✓ EtO Sterilized (SAL 10^{-6})	Pass
✓ Precise pH interpretation, pHID (when used at 35°-38°C)	**B3435**

DEKRA certified — Quality — ISO 13485 ISO 9001 — www.dekra-seal.com

Steve Geelhood
Business Director

04 Feb 2022
Date

BCSI Internal Tracking: 2202006

400003.06

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI. THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

BCSI