



Blood Cell Storage, Inc.  
4203 23rd Ave W,  
Seattle WA, 98199 USA

# Certificate of Compliance

Date of EtO Processing: 13 APR 2022

Product Name: SAFE Sens<sup>®</sup> sv<sup>2</sup> sensor kit

Part Number: 400003

Lot Number: 2203047

Storage: Room temperature protected from light


EXPIRY: 30 November 2024

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- ✓ Mouse embryo toxicity Pass
- ✓ Biological Indicator Culturing Test Pass
- ✓ EtO Sterilized (SAL 10<sup>-6</sup>) Pass
- ✓ Precise pH interpretation, pHID **B0288**  
(when used at 35°-38°C)



  
Steve Geelhood  
Business Director

10 Mar 23  
Date

BCSI Internal Tracking: 2303016

400003.06

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI.  
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN  
PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

