



Blood Cell Storage, Inc.
4203 23rd Ave W
Seattle WA, 98199 USA

Certificate of Compliance

Date of EtO Processing: 06 February 2023

Product Name: SAFE Sens[®] sv² sensor kit

Part Number: 400009

Lot Number: 2210009

Storage: Room temperature protected from light

EXPIRY: 31 December 2025

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- | | |
|--|--------------|
| ✓ Mouse embryo toxicity | Pass |
| ✓ Biological Indicator Culturing Test | Pass |
| ✓ EtO Sterilized (SAL 10 ⁻⁶) | Pass |
| ✓ Precise pH interpretation, encrypted pHID
(when used at 35°-38°C) | 45TU% |



DocuSigned by Hannah Erickson
Hannah Erickson | I approve this document
11 January 2024 | 1:13:53 PM PST
B77B1CEF28B346A1A35393D78EF4DE29

11 January 2024

Hannah Erickson
Quality Assurance

Date

BCSI Internal Tracking: 2401019

400009.06

**THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI.
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN
PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.**

