

Date of Manufacture: 26 August, 2020

Storage: Room temperature protected from light

BCSI certifies that the above referenced product:

Product Name: sv² sensor kit

EXPIRY: 28 February, 2022

Part Number: 400019

Lot Number: 1911001

BCSI

Blood Cell Storage, Inc. 123 NW 36th Street, Suite 220 Seattle, WA 98107 USA

> Quality ISO 13485 ISO 9001

Certificate of Compliance

accredited registering bod	se Quality Management System is approved by an ly to the ISO 9001:2015 Quality Systems Standard lical Devices Quality Management System.
 Was tested and complies 	with applicable BCSI Standard Procedures. with visual and performance I in the product specification, including:
✓ Mouse embryo tox	icity Pass
✓ Non-pyrogenic (La	AL test) Pass
✓ EtO Sterilized (SA	L 10 ⁻⁶) Pass
✓ Precise pH interpre (when used at 35°-	etation, encrypted pHID 0IR6S -38°C)
the	28 Ang 2000
Irena Dikovsky	Date
Director of Quality	
BCSI Reference: 2008026	Rev 400019.01

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI. THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.