



Certificate of Compliance

Date of EtO Processing: 21 August 2020	
Product Name: sv ² sensor kit	
Part Number: 400019	
Lot Number: 2008003	
Storage: Room temperature protected from ligh	t
EXPIRY: <u>31 March 2023</u>	
 Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard and ISO 13485:2016 Medical Devices Quality Management System. Was tested in accordance with applicable BCSI Standard Procedures. Was tested and complies with visual and performance specifications as specified in the product specification, including: ✓ Mouse embryo toxicity ✓ Biological Indicator Culturing Test ✓ EtO Sterilized (SAL 10⁻⁶) ✓ Pass ✓ Precise pH interpretation, encrypted pHID (when used at 35°-38°C) 	
Irena Dikovsky Director of Quality	Date
BCSI Reference: 2104009	Rev 400019.02

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI. THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

