



# Certificate of Compliance

Date of EtO Processing: 26 March 2021

Product Name: sv<sup>2</sup> sensor kit

Part Number: 400019

Lot Number: 2103001

Storage: Room temperature protected from light

EXPIRY: 30 April 2023

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard and ISO 13485:2016 Medical Devices Quality Management System.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- |  |              |
|--|--------------|
| ✓ Mouse embryo toxicity  | Pass         |
| ✓ Biological Indicator Culturing Test                                  | Pass         |
| ✓ EtO Sterilized (SAL 10 <sup>-6</sup> )                               | Pass         |
| ✓ Precise pH interpretation, encrypted pHID<br>(when used at 35°-38°C) | <b>5@NTZ</b> |



Irena Dikovskiy  
Director of Quality

Date

BCSI Reference: 2104043

Rev 400019.02

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI.  
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN  
PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

