



Certificate of Compliance

Date of EtO Processing: 06 February 2023

Product Name: sv² sensor kit

Part Number: 400019

Lot Number: 2208028

Storage: Room temperature protected from light

EXPIRY: 31 May 2025

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the ISO 9001:2015 Quality Systems Standard and ISO 13485:2016 Medical Devices Quality Management System.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification, including:

- | | |
|--|--------------|
| ✓ Mouse embryo toxicity | Pass |
| ✓ Biological Indicator Culturing Test | Pass |
| ✓ EtO Sterilized (SAL 10 ⁻⁶) | Pass |
| ✓ Precise pH interpretation, encrypted pHID
(when used at 35°-38°C) | 0MZ6S |



DocuSigned by Hannah Erickson
 | I approve this document
20 September 2023 | 8:48:58 AM PDT
B77B1CECF28B346A1A35393D78EF4DE29

Hannah Erickson
Quality Assurance

20 September 2023

Date

BCSI Reference: 2309009

Rev 400019.04

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI. THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM BCSI'S QUALITY ASSURANCE MANAGER.

