



Blood Cell Storage, Inc.
 4203 23rd Ave W.
 Seattle WA, 98199 USA

Certificate of Compliance

Date of Manufacture: 17 June 2022

Product Name: BCSI pH SAFE Reader (pH1000™)

Part Number: 400001

Lot Number: 2206010[†]

BCSI pH1000 Serial Number: BCS0809P0110

BCSI pH1000 QC Standard Serial Number: B397

Serial Number	Fr	Fp	Rp
B397	1602	2574	1.729



B3971602257417290141

NOTE: These criteria are specific for the pH1000 paired with the QC Standard as stated above. The Fr and Fp values must be within 5% of the listed value. The Rp value must be within 1% of the listed value.

USER: _____ EXPIRY[†]: _____
 Date QC Standard implemented (barcode scanned) One (1) year from date of barcode scan

USER: Record the expiry date on the SQC label.

BCSI certifies that the above referenced product:

- Was tested by BCSI whose Quality Management System is approved by an accredited registering body to the appropriate ISO 13485:2016 Medical Devices Quality Systems Standard and ISO 9001:2015 Quality Systems Standard.
- Was tested in accordance with applicable BCSI Standard Procedures.
- Was tested and complies with visual and performance specifications as specified in the product specification.



Steve Geelhood
 Business Director

17 June 2022
 Date

400001-07

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF BCSI. THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM BCSI'S QUALITY ASSURANCE DEPARTMENT.